



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:

Muncy Industries, LLC.
5820 Susquehanna Trail, Turbotville, PA 17772

*and hereby declares that the Organization is accredited in accordance with
the recognized International Standard:*

**ISO/IEC 17025:2017
& Meets the Requirements of ANSI/NSCL Z540.1-1994**

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

Mass, Force, and Weighing Devices Calibration
(As detailed in the supplement)

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope. This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

Initial Accreditation Date:

Issue Date:

Expiration Date:

April 17, 2019

December 31, 2025

January 31, 2028

Accreditation No.:

Certificate No.:

97822

L25-1009

Tracy Szerszen
President

*The validity of this certificate is maintained through ongoing assessments based
on a continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjilabs.com*

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084



Certificate of Accreditation: Supplement

Muncy Industries, LLC.
 5820 Susquehanna Trail, Turbotville, PA 17772
 Contact Name: Shawn Millard Phone: 570-649-5188

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY (\pm) ¹	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Mass, Force, and Weighing Devices	Equipment to Source and Measure Force (Compression and Tension)	500 lbf to 5 000 lbf	0.81 % of reading	TECSIS 50 000 lbf loadcell	ASTM E-4	F1, F2	O
Mass, Force, and Weighing Devices	Equipment to Source and Measure Force (Compression and Tension)	5 000 lbf to 50 000 lbf	0.18 % of reading	TECSIS 50 000 lbf loadcell	ASTM E-4	F1, F2	O
Mass, Force, and Weighing Devices	Equipment to Source and Measure Force (Compression and Tension)	2 500 lbf to 25 000 lbf	0.59 % of reading	TECSIS 250 000 lbf loadcell	ASTM E-4	F1, F2	O
Mass, Force, and Weighing Devices	Equipment to Source and Measure Force (Compression and Tension)	25 000 lbf to 250 000 lbf	0.21 % of reading	TECSIS 250 000 lbf loadcell	ASTM E-4	F1, F2	O
Mass, Force, and Weighing Devices	Equipment to Source and Measure Force (Compression and Tension)	6 500 lbf to 65 000 lbf	0.58 % of reading	STRAINSENSE 650 000 lbf loadcell	ASTM E-4	F1, F2	O
Mass, Force, and Weighing Devices	Equipment to Source and Measure Force (Compression and Tension)	65 000 lbf to 650 000 lbf	0.31 % of reading	STRAINSENSE 650 000 lbf loadcell	ASTM E-4	F1, F2	O
Mass, Force, and Weighing Devices	Equipment to Source and Measure Force (Compression and Tension)	13 000 lbf to 130 000 lbf	0.44 % of reading	Two STRAINSENSE 650 000 lbf loadcells	ASTM E-4	F1, F2	O
Mass, Force, and Weighing Devices	Equipment to Source and Measure Force (Compression and Tension)	130 000 lbf to 1 300 000 lbf	0.59 % of reading	Two STRAINSENSE 650 000 lbf loadcells	ASTM E-4	F1, F2	O



Certificate of Accreditation: Supplement

Muncy Industries, LLC.
5820 Susquehanna Trail, Turbotville, PA 17772
Contact Name: Shawn Millard Phone: 570-649-5188

Accreditation is granted to the facility to perform the following conformity assessment activities:

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. Location of activity:

Location Code	Location
O	Conformity assessment activity is performed onsite at the CABs customer location
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.